

We, ABB Limited

Of, Oldends Lane Stonehouse Gloucestershire England GL10 3TA

Declare under our own responsibility that the product: Gas Analyser system Type 6553 (Certificate BAS01ATEX7043) when fitted with MTL7755ac barrier. to which this declaration relates is in conformity with the following standards:

Directive 94/9/EC (ATEX)

Harmonised Standard Used: EN 60079-0:2011+A11:2013 EN 60079-11:2011 EN60079-25:2010

Provisions of the Directive fulfilled by the Equipment. Equipment Group:-



Ex ia IIC T4

Name, notified body number and address of the Notified Body:-Baseefa Ltd, Number 1180, of:- Rockhead Business Park, Staden Lane, Buxton, Derbyshire SK17 9RZ

Number of the type examination certificate:Ex 01E2044 - last supplement issued 15th January 2016

When installed and operated according to the supplied instructions.

12/10/16

Date of Issue

Signature Y.Saulick
Certification Engineer